

# Manual Supplement

Manual Title: 345 Users	Supplement Issue: <b>2</b>
Part Number: 2560401	Issue Date: 3/08
Print Date: October 2006	Page Count: 1
Revision/Date:	

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 15, under **Voltage Measuring Input**, replace the first sentence with the following:

The maximum input voltage for over-voltage category CAT IV may not exceed 600 V to earth ground.

## Change #2

On page 25, under **Balanced Three-Phase Power Measurement Connection**, replace:

Current is measured on one phase and the two voltages are measured on the remaining phase.

With:

Current is measured on one phase and the phase-to-phase voltage is measured on the remaining line.

On page 55, replace the screen shot with the following:

Power 1 phase			2006-03-29, 09:32
		50.0 Hz	
kW	kVA	kVAR	
2.34	4.62	3.98	
V rms	A rms	PF	
230.0	20.06	0.507	